



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calipers <sup>FO</sup>	0.05 in to 40 in	(230 + 9.3L) $\mu$ in	Gage Blocks IC-DM-2
Height Gages <sup>FO</sup>	Up to 40 in	(570 + 23L) $\mu$ in	Gage Blocks IC-DM-3
Gage Blocks <sup>F</sup>	0.05 in to 4 in	(2 + 1.1L) $\mu$ in	Gage Block Comparator IC-DM-1
	0.05 in to 14 in	(3 + 1.1L) $\mu$ in	P&W Labmaster IC-DM-1
Dial/Digital/Test Indicators <sup>FO</sup>	Up to 6 in	(55 + 24L) $\mu$ in	Gage Blocks IC-DM-22
Cylindrical Plug Gages <sup>FO</sup>	Up to 6 in	(6.6 + 2.6L) $\mu$ in	P&W Labmaster IC-DM-15
Plain Ring Gages <sup>F</sup>	0.02 in to 14 in	(15 + 2.9L) $\mu$ in	P&W Labmaster IC-DM-8
Outside Micrometers <sup>FO</sup>	0.05 in to 40 in	(55 + 18L) $\mu$ in	Gage Blocks IC-DM-2
Inside Micrometers <sup>FO</sup>	0.05 in to 40 in	(66 + 18L) $\mu$ in	Gage Blocks IC-DM-29
Depth Micrometers <sup>FO</sup>	0.05 in to 12 in	(66 + 25L) $\mu$ in	Gage Blocks IC-DM-26
Steel Rules <sup>FO</sup>	Up to 40 in	(1 200 + 110L) $\mu$ in	Linear Calibrator IC-DM-31
Tape Measures <sup>FO</sup>	Up to 40 in	(1 200 + 110L) $\mu$ in	Linear Calibrator IC-DM-28
Feeler Gages <sup>FO</sup>	0.001 in to 0.2 in	16 $\mu$ in	Supermicrometer IC-DM-27



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Voltage <sup>FO</sup>	0.01 mV to 330 mV	20 $\mu$ V/V + 1 $\mu$ V	Multifunction Calibrator OEM Manual
	330 mV to 3.3 V	11 $\mu$ V/V + 2 $\mu$ V	
	3.3 V to 33 V	12 $\mu$ V/V + 20 $\mu$ V	
	33 V to 330 V	18 $\mu$ V/V + 150 $\mu$ V	
	330 V to 1 000 V	18 $\mu$ V/V + 1.5 mV	
Equipment to Output DC Voltage <sup>FO</sup>	0.01 mV to 100 mV	11 $\mu$ V/V + 3 $\mu$ V/V	Multifunction Calibrator OEM Manual
	100 mV to 1 V	10 $\mu$ V/V + 0.3 $\mu$ V/V	
	1 V to 10 V	10 $\mu$ V/V + 0.1 $\mu$ V/V	
	10 V to 100 V	12 $\mu$ V/V + 2.3 $\mu$ V/V	
	100 V to 1 000 V	24 $\mu$ V/V + 0.1 $\mu$ V/V	
Equipment to Measure Capacitance <sup>FO</sup>	220 pF to 400 pF	10 pF + 0.5% of Reading	Multifunction Calibrator OEM Manual
	400 pF to 1.1 nF	0.01 nF + 0.5% of Reading	
	1.1 nF to 3.3 nF	0.01 nF + 0.5% of Reading	
	3.3 nF to 11 nF	0.01 nF + 0.25% of Reading	
	11 nF to 33 nF	0.1 nF + 0.25% of Reading	
	33 nF to 110 nF	0.1 nF + 0.25% of Reading	
	110 nF to 330 nF	0.3 nF + 0.25% of Reading	
	330 nF to 1.1 $\mu$ F	1 nF + 0.25% of Reading	
	1.1 $\mu$ F to 3.3 $\mu$ F	3 nF + 0.25% of Reading	
	3.3 $\mu$ F to 11 $\mu$ F	10 nF + 0.25% of Reading	
	11 $\mu$ F to 33 $\mu$ F	30 nF + 0.4% of Reading	
	33 $\mu$ F to 110 $\mu$ F	100 nF + 0.45% of Reading	
	110 $\mu$ F to 330 $\mu$ F	300 nF + 0.45% of Reading	
	330 $\mu$ F to 1.1 mF	1 $\mu$ F + 0.45% of Reading	
	1.1 mF to 3.3 mF	3 $\mu$ F + 0.45% of Reading	
3.3 mF to 11 mF	10 $\mu$ F + 0.45% of Reading		
11 mF to 33 mF	30 $\mu$ F + 0.75% of Reading		
33 mF to 110 mF	100 $\mu$ F + 1.1% of Reading		



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Current <sup>FO</sup>	330 $\mu$ A to 3.3 mA	105 $\mu$ A/A + 0.05 $\mu$ A	Multifunction Calibrator OEM Manual
	3.3 mA to 33 mA	100 $\mu$ A/A + 0.2 $\mu$ A	
	33 mA to 330 mA	100 $\mu$ A/A + 2.5 $\mu$ A	
	330 mA to 1.1 A	200 $\mu$ A/A + 40 $\mu$ A	
	1.1 A to 3 A	380 $\mu$ A/A + 40 $\mu$ A	
	3 A to 11 A	500 $\mu$ A/A + 500 $\mu$ A	
	11 A to 20.5 A	1 000 $\mu$ A/A + 750 $\mu$ A	
Equipment to Output DC Current <sup>FO</sup>	0.01 nA to 100 nA	35 $\mu$ A/A + 400 $\mu$ A/A	8.5 Digit Multimeter OEM Manual
	1 nA to 1 $\mu$ A	25 $\mu$ A/A + 40 $\mu$ A/A	
	1 $\mu$ A to 10 $\mu$ A	25 $\mu$ A/A + 10 $\mu$ A/A	
	10 $\mu$ A to 100 $\mu$ A	25 $\mu$ A/A + 8 $\mu$ A/A	
	100 $\mu$ A to 1 mA	25 $\mu$ A/A + 5 $\mu$ A/A	
	1 mA to 10 mA	25 $\mu$ A/A + 5 $\mu$ A/A	
	10 mA to 100 mA	40 $\mu$ A/A + 5 $\mu$ A/A	
	100 mA to 1 A	115 $\mu$ A/A + 10 $\mu$ A/A	
1.1 A to 20 A	0.02 % of Reading		
Equipment to Measure DC High Voltage <sup>FO</sup>	0 kV to 30 kV	0.01 % of Reading + 0.01 % of Full Scale	Vitretek 4700 HVL-35 Probe OEM Manual
Equipment to Measure DC Power <sup>FO</sup>			Multifunction Calibrator OEM Manual
0.33 mA to 330 mA	33 mV to 1 020 V	0.03 % of Reading	
0.33 A to 3 A	33 mV to 1 020 V	0.03 % of Reading	
3 A to 20.5 A	33 mV to 1 020 V	0.07 % of Reading	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			Multifunction Calibrator OEM Manual
10 Hz to 20 Hz	29 $\mu$ A to 330 $\mu$ A	0.2 % of Reading + 0.1 $\mu$ A	
20 Hz to 45 Hz	29 $\mu$ A to 330 $\mu$ A	0.15 % of Reading + 0.1 $\mu$ A	
45 Hz to 1 kHz	29 $\mu$ A to 330 $\mu$ A	0.13 % of Reading + 0.1 $\mu$ A	
5 kHz to 10 kHz	29 $\mu$ A to 330 $\mu$ A	0.8 % of Reading + 0.2 $\mu$ A	
10 kHz to 30 kHz	29 $\mu$ A to 330 $\mu$ A	1.6 % of Reading + 0.4 $\mu$ A	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			Multifunction Calibrator OEM Manual
10 Hz to 20 Hz	0.33 mA to 3.3 mA	0.2 % of Reading + 0.15 $\mu$ A	
20 Hz to 45 Hz	0.33 mA to 3.3 mA	0.13 % of Reading + 0.15 $\mu$ A	
45 Hz to 1 kHz	0.33 mA to 3.3 mA	0.1 % of Reading + 0.15 $\mu$ A	
1 kHz to 5 kHz	0.33 mA to 3.3 mA	0.2 % of Reading + 0.2 $\mu$ A	
5 kHz to 10 kHz	0.33 mA to 3.3 mA	0.5 % of Reading + 0.3 $\mu$ A	
10 kHz to 30 kHz	0.33 mA to 3.3 mA	1 % of Reading + 0.6 $\mu$ A	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			
10 Hz to 20 Hz	3.3 mA to 33 mA	0.18 % of Reading + 2 $\mu$ A	
20 Hz to 45 Hz	3.3 mA to 33 mA	0.09 % of Reading + 2 $\mu$ A	
45 Hz to 1 kHz	3.3 mA to 33 mA	0.04 % of Reading + 2 $\mu$ A	
1 kHz to 5 kHz	3.3 mA to 33 mA	0.08 % of Reading + 2 $\mu$ A	
5 kHz to 10 kHz	3.3 mA to 33 mA	0.2 % of Reading + 3 $\mu$ A	
10 kHz to 30 kHz	3.3 mA to 33 mA	0.4 % of Reading + 4 $\mu$ A	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			
10 Hz to 20 Hz	33 mA to 330 mA	0.18 % of Reading + 20 $\mu$ A	
20 Hz to 45 Hz	33 mA to 330 mA	0.09 % of Reading + 20 $\mu$ A	
45 Hz to 1 kHz	33 mA to 330 mA	0.04 % of Reading + 20 $\mu$ A	
1 kHz to 5 kHz	33 mA to 330 mA	0.1 % of Reading + 50 $\mu$ A	
5 kHz to 10 kHz	33 mA to 330 mA	0.2 % of Reading + 100 $\mu$ A	
10 kHz to 30 kHz	33 mA to 330 mA	0.4 % of Reading + 200 $\mu$ A	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	0.33 A to 1.1 A	0.18 % of Reading + 100 $\mu$ A	
45 Hz to 1 kHz	0.33 A to 1.1 A	0.05 % of Reading + 100 $\mu$ A	
1 kHz to 5 kHz	0.33 A to 1.1 A	0.6 % of Reading + 1 000 $\mu$ A	
5 kHz to 10 kHz	0.33 A to 1.1 A	2.5 % of Reading + 5 000 $\mu$ A	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	1.1 A to 3 A	0.18 % of Reading + 100 $\mu$ A	
45 Hz to 1 kHz	1.1 A to 3 A	0.05 % of Reading + 100 $\mu$ A	
1 kHz to 5 kHz	1.1 A to 3 A	0.6 % of Reading + 1 000 $\mu$ A	
5 kHz to 10 kHz	1.1 A to 3 A	2.5 % of Reading + 5 000 $\mu$ A	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			Multifunction Calibrator OEM Manual
45 Hz to 100 Hz	3 A to 11 A	0.06 % of Reading + 2 000 $\mu$ A	
100 Hz to 1 kHz	3 A to 11 A	0.1 % of Reading + 2 000 $\mu$ A	
1 kHz to 5 kHz	3 A to 11 A	3 % of Reading + 2 000 $\mu$ A	
Equipment to Measure AC Current (At the listed frequencies) <sup>FO</sup>			
45 Hz to 100 Hz	11 A to 20.5 A	0.12 % of Reading + 5 000 $\mu$ A	
100 Hz to 1 kHz	11 A to 20.5 A	0.15 % of Reading + 5 000 $\mu$ A	
1 kHz to 5 kHz	11 A to 20.5 A	3 % of Reading + 5 000 $\mu$ A	
Equipment to Output AC Current (At the listed frequencies) <sup>FO</sup>			
10 Hz to 20 Hz	100 mA to 1 A	0.4 % of Full Scale + 0.02 % of Reading	
20 Hz to 45 Hz	100 mA to 1 A	0.16 % of Full Scale + 0.02 % of Reading	
45 Hz to 100 Hz	100 mA to 1 A	0.08 % of Full Scale + 0.02 % of Reading	
100 Hz to 5 kHz	100 mA to 1 A	0.1 % of Full Scale + 0.02 % of Reading	
5 kHz to 20 kHz	100 mA to 1 A	0.03 % of Full Scale + 0.02 % of Reading	
20 kHz to 50 kHz	100 mA to 1 A	1 % of Full Scale + 0.04 % of Reading	
Equipment to Output AC Current (At the listed frequencies) <sup>FO</sup>			
45 Hz to 5 kHz	1 A to 20 A	0.15 % of Reading + 16 mA	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	1 mV to 33 mV	800 $\mu$ V/V + 6 $\mu$ V	
45 Hz to 10 kHz	1 mV to 33 mV	150 $\mu$ V/V + 6 $\mu$ V	
10 kHz to 20 kHz	1 mV to 33 mV	200 $\mu$ V/V + 6 $\mu$ V	
20 kHz to 50 kHz	1 mV to 33 mV	1 000 $\mu$ V/V 6 $\mu$ V	
50 kHz to 100 kHz	1 mV to 33 mV	3 500 $\mu$ V/V 12 $\mu$ V	
100 kHz to 500 kHz	1 mV to 33 mV	8 000 $\mu$ V/V + 50 $\mu$ V	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	33 mV to 330 mV	300 $\mu$ V/V + 8 $\mu$ V	
50 kHz to 100 kHz	33 mV to 330 mV	800 $\mu$ V/V + 32 $\mu$ V	
100 kHz to 500 kHz	33 mV to 330 mV	2 000 $\mu$ V/V + 70 $\mu$ V	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			Multifunction Calibrator OEM Manual
10 Hz to 45 Hz	0.33 V to 3.3 V	300 $\mu$ V/V + 50 $\mu$ V	
20 kHz to 50 kHz	0.33 V to 3.3 V	300 $\mu$ V/V + 50 $\mu$ V	
50 kHz to 100 kHz	0.33 V to 3.3 V	700 $\mu$ V/V + 125 $\mu$ V	
100 kHz to 500 kHz	0.33 V to 3.3 V	2 400 $\mu$ V/V + 600 $\mu$ V	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	3.3 V to 33 V	300 $\mu$ V/V + 650 $\mu$ V	
10 kHz to 20 kHz	3.3 V to 33 V	240 $\mu$ V/V + 600 $\mu$ V	
20 kHz to 50 kHz	3.3 V to 33 V	350 $\mu$ V/V + 600 $\mu$ V	
50 kHz to 100 kHz	3.3 V to 33 V	900 $\mu$ V/V + 1 600 $\mu$ V	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	1 mV to 33 mV	0.08 % of Reading + 6 $\mu$ V	
45 Hz to 10 kHz	1 mV to 33 mV	0.015 % of Reading + 6 $\mu$ V	
10 kHz to 20 kHz	1 mV to 33 mV	0.02 % of Reading + 6 $\mu$ V	
20 kHz to 50 kHz	1 mV to 33 mV	0.1 % of Reading + 6 $\mu$ V	
50 kHz to 100 kHz	1 mV to 33 mV	0.35 % of Reading + 12 $\mu$ V	
100 kHz to 500 kHz	1 mV to 33 mV	0.8 % of Reading + 50 $\mu$ V	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	33 mV to 330 mV	0.03 % of Reading + 8 $\mu$ V	
45 Hz to 10 kHz	33 mV to 330 mV	0.013 % of Reading + 8 $\mu$ V	
10 kHz to 20 kHz	33 mV to 330 mV	0.015 % of Reading + 8 $\mu$ V	
20 kHz to 50 kHz	33 mV to 330 mV	0.035 % of Reading + 8 $\mu$ V	
50 kHz to 100 kHz	33 mV to 330 mV	0.08 % of Reading + 32 $\mu$ V	
100 kHz to 500 kHz	33 mV to 330 mV	0.2 % of Reading + 70 $\mu$ V	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>FO</sup>			
10 Hz to 45 Hz	0.33 V to 3.3 V	0.03 % of Reading + 50 $\mu$ V	
45 Hz to 10 kHz	0.33 V to 3.3 V	0.012 % of Reading + 25 $\mu$ V	
10 kHz to 20 kHz	0.33 V to 3.3 V	0.019 % of Reading + 50 $\mu$ V	
20 kHz to 50 kHz	0.33 V to 3.3 V	0.03 % of Reading + 50 $\mu$ V	
50 kHz to 100 kHz	0.33 V to 3.3 V	0.07 % of Reading + 125 $\mu$ V	
100 kHz to 500 kHz	0.33 V to 3.3 V	0.24 % of Reading + 600 $\mu$ V	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage (At the listed frequencies) <sup>F0</sup>			Multifunction Calibrator OEM Manual
10 Hz to 45 Hz	3.3 V to 33 V	0.03 % of Reading + 650 $\mu$ V	
45 Hz to 10 kHz	3.3 V to 33 V	0.015 % of Reading + 200 $\mu$ V	
10 kHz to 20 kHz	3.3 V to 33 V	0.024 % of Reading + 600 $\mu$ V	
20 kHz to 50 kHz	3.3 V to 33 V	0.035 % of Reading + 600 $\mu$ V	
50 kHz to 100 kHz	3.3 V to 33 V	0.09 % of Reading + 1.6 mV	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>F0</sup>			
45 Hz to 1 kHz	33 V to 330 V	0.019 % of Reading + 2 mV	
1 kHz to 10 kHz	33 V to 330 V	0.02 % of Reading + 6 mV	
10 kHz to 20 kHz	33 V to 330 V	0.025 % of Reading + 6 mV	
20 kHz to 50 kHz	33 V to 330 V	0.03 % of Reading + 6 mV	
50 kHz to 100 kHz	33 V to 330 V	0.2 % of Reading + 50 mV	
Equipment to Measure AC Voltage (At the listed frequencies) <sup>F0</sup>			
45 Hz to 1 kHz	330 V to 1 020 V	0.03 % of Reading + 10 mV	
1 kHz to 5 kHz	330 V to 1 020 V	0.025 % of Reading + 10 mV	
5 kHz to 10 kHz	330 V to 1 020 V	0.03 % of Reading + 10 mV	
Equipment to Measure AC Voltage -Bandwidth < 2 MHz (At the listed frequencies) <sup>F0</sup>			Multifunction Calibrator OEM Manual
1 Hz to 40 Hz	1 mV to 10 mV	0.03 % of Reading + 3 $\mu$ V	
40 Hz to 1 kHz	1 mV to 10 mV	0.02 % of Reading + 1.1 $\mu$ V	
1 kHz to 20 kHz	1 mV to 10 mV	0.03 % of Reading + 1.1 $\mu$ V	
20 kHz to 50 kHz	1 mV to 10 mV	0.1 % of Reading + 1.1 $\mu$ V	
50 kHz to 100 kHz	1 mV to 10 mV	0.5 % of Reading + 1.1 $\mu$ V	
100 kHz to 300 kHz	1 mV to 10 mV	4 % of Reading + 2 $\mu$ V	
Equipment to Measure AC Voltage -Bandwidth < 2 MHz (At the listed frequencies) <sup>F0</sup>			
1 Hz to 40 Hz	100 mV to 10 V	0.007 % of Reading + 0.4 mV	
40 Hz to 1 kHz	100 mV to 10 V	0.007 % of Reading + 0.2 mV	
1 kHz to 20 kHz	100 mV to 10 V	0.014 % of Reading + 0.2 mV	
20 kHz to 50 kHz	100 mV to 10 V	0.03 % of Reading + 0.2 mV	
50 kHz to 100 kHz	100 mV to 10 V	0.08 % of Reading + 0.2 mV	
100 kHz to 300 kHz	100 mV to 10 V	0.3 % of Reading + 1 mV	
300 kHz to 1 MHz	100 mV to 10 V	1 % of Reading + 1 mV	
1 MHz to 2 MHz	100 mV to 10 V	1.5 % of Reading + 1 mV	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage -Bandwidth < 2 MHz (At the listed frequencies) <sup>FO</sup>			Multifunction Calibrator OEM Manual
1 Hz to 40 Hz	10 V to 100 V	0.02 % of Reading + 4 mV	
40 Hz to 1 kHz	10 V to 100 V	0.02 % of Reading + 2 mV	
1 kHz to 20 kHz	10 V to 100 V	0.02 % of Reading + 2 mV	
20 kHz to 50 kHz	10 V to 100 V	0.035 % of Reading + 2 mV	
50 kHz to 100 kHz	10 V to 100 V	0.12 % of Reading + 2 mV	
100 kHz to 300 kHz	10 V to 100 V	0.4 % of Reading + 10 mV	
300 kHz to 1 MHz	10 V to 100 V	1.5 % of Reading + 10 mV	
Equipment to Measure AC Voltage -Bandwidth < 2 MHz (At the listed frequencies) <sup>FO</sup>			
1 Hz to 40 Hz	100 V to 1 000 V	0.04 % of Reading + 40 mV	
40 Hz to 1 kHz	100 V to 1 000 V	0.04 % of Reading + 20 mV	
1 kHz to 20 kHz	100 V to 1 000 V	0.06 % of Reading + 20 mV	
20 kHz to 50 kHz	100 V to 1 000 V	0.12 % of Reading + 20 mV	
50 kHz to 100 kHz	100 V to 1 000 V	0.3 % of Reading + 20 mV	
Equipment to Measure AC Voltage -Bandwidth > 2 MHz (At the listed frequencies) <sup>FO</sup>			
45 Hz to 100 kHz	0.01 mV to 10 mV	0.09 % of Reading + 6 $\mu$ V	
100 kHz to 1 MHz	0.01 mV to 10 mV	1.2 % of Reading + 5 $\mu$ V	
1 MHz to 4 MHz	0.01 mV to 10 mV	7 % of Reading + 7 $\mu$ V	
4 MHz to 8 MHz	0.01 mV to 10 mV	20 % of Reading + 8 $\mu$ V	
Equipment to Measure AC Voltage -Bandwidth > 2 MHz (At the listed frequencies) <sup>FO</sup>			
45 Hz to 100 kHz	100 mV to 10 V	0.09 % of Reading + 6 mV	
100 kHz to 1 MHz	100 mV to 10 V	2 % of Reading + 5 mV	
1 MHz to 4 MHz	100 mV to 10 V	4 % of Reading + 7 mV	
4 MHz to 8 MHz	100 mV to 10 V	4 % of Reading + 8 mV	
8 MHz to 10 MHz	100 mV to 10 V	15 % of Reading + 10 mV	
Equipment to Measure AC Voltage -Bandwidth > 2 MHz (At the listed frequencies) <sup>FO</sup>			
45 Hz to 100 kHz	10 V to 100 V	0.12 % of Reading + 2 mV	
Equipment to Measure AC Voltage -Bandwidth > 2 MHz (At the listed frequencies) <sup>FO</sup>			
45 Hz to 100 kHz	100 V to 1 000 V	0.3 % of Reading + 0.1 V	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output Frequency <sup>FO</sup>	0.01 Hz to 2 MHz	2.5 Hz/MHz + 5 $\mu$ Hz	8.5 Digit Multimeter OEM Manual
Equipment to Measure Frequency <sup>FO</sup>	1 Hz to 10 MHz	0.05 % of Reading	Multifunction Calibrator OEM Manual
	1 Hz to 40 Hz	0.05 % of Reading	
	40 Hz to 10 MHz	0.01 % of Reading	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type B <sup>FO</sup>	600 °C to 800 °C	0.44 °C	Electrical Simulation of Thermocouple Output OEM Manual
	800 °C to 1 000 °C	0.34 °C	
	1 000 °C to 1 550 °C	0.3 °C	
	1 550 °C to 1 820 °C	0.33 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type C <sup>FO</sup>	0 °C to 150 °C	0.3 °C	
	150 °C to 650 °C	0.26 °C	
	650 °C to 1 000 °C	0.31 °C	
	1 000 °C to 1 800 °C	0.5 °C	
	1 800 °C to 2 316 °C	0.84 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type E <sup>FO</sup>	-250 °C to -100 °C	0.5 °C	
	-100 °C to -25 °C	0.16 °C	
	-25 °C to 350 °C	0.14 °C	
	350 °C to 650 °C	0.16 °C	
	650 °C to 1 000 °C	0.21 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type J <sup>FO</sup>	-210 °C to -100 °C	0.27 °C	
	-100 °C to -30 °C	0.16 °C	
	-30 °C to 150 °C	0.14 °C	
	150 °C to 760 °C	0.17 °C	
	760 °C to 1 200 °C	0.23 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type K <sup>FO</sup>	-200 °C to -100 °C	0.33 °C	
	-100 °C to -25 °C	0.18 °C	
	-25 °C to 120 °C	0.16 °C	
	120 °C to 1 000 °C	0.26 °C	
	1 000 °C to 1 372 °C	0.4 °C	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication and Control Equipment Used with Thermocouple Type L <sup>FO</sup>	-200 °C to -100 °C	0.37 °C	Electrical Simulation of Thermocouple Output OEM Manual
	-100 °C to 800 °C	0.26 °C	
	800 °C to 900 °C	0.17 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type N <sup>FO</sup>	-200 °C to -100 °C	0.4 °C	
	-100 °C to -25 °C	0.22 °C	
	-25 °C to 120 °C	0.19 °C	
	120 °C to 410 °C	0.18 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type R <sup>FO</sup>	410 °C to 1 300 °C	0.27 °C	
	0 °C to 250 °C	0.57 °C	
	250 °C to 400 °C	0.35 °C	
	400 °C to 1 000 °C	0.33 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type S <sup>FO</sup>	1 000 °C to 1 767 °C	0.4 °C	
	0 °C to 250 °C	0.47 °C	
	250 °C to 1 000 °C	0.36 °C	
	1 000 °C to 1 400 °C	0.37 °C	
Temperature Calibration, Indication and control Equipment used with Thermocouple Type T <sup>FO</sup>	1 400 °C to 1 767 °C	0.46 °C	
	-250 °C to -150 °C	0.63 °C	
	-150 °C to 0 °C	0.24 °C	
	0 °C to 120 °C	0.16 °C	
Temperature Calibration, Indication and Control Equipment Used with Thermocouple Type U <sup>FO</sup>	120 °C to 400 °C	0.14 °C	
	-200 °C to 0 °C	0.56 °C	
Temperature Calibration, Indication and Control Equipment Used with Thermocouple Type U <sup>FO</sup>	0 °C to 600 °C	0.27 °C	
	-200 °C to 800 °C	0.23 °C	Electrical Simulation of RTD Output OEM Manual
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 395, 100 $\Omega$ <sup>FO</sup>			
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 3926, 100 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.12 °C	
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 3916, 100 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.25 °C	



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 385, 200 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.16 °C	Electrical Simulation of RTD Output OEM Manual
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 385, 500 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.11 °C	
Temperature Calibration, Indication and control Equipment used with RTD Type Pt 385, 1 000 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.23 °C	
Temperature Calibration, Indication and control Equipment used with RTD Type Pt Ni 385, 120 $\Omega$ <sup>FO</sup>	-80 °C to 260 °C	0.14 °C	
Temperature Calibration, Indication and control Equipment used with RTD Type Cu 427, 10 $\Omega$ <sup>FO</sup>	-100 °C to 260 °C	0.3 °C	
Equipment to Measure Resistance <sup>FO</sup>	Up to 11 $\Omega$	40 $\mu\Omega/\Omega$ + 1 m $\Omega$	Multifunction Calibrator OEM Manual
	11 $\Omega$ to 33 $\Omega$	30 $\mu\Omega/\Omega$ + 1.5 m $\Omega$	
	33 $\Omega$ to 110 $\Omega$	28 $\mu\Omega/\Omega$ + 1.4 m $\Omega$	
	110 $\Omega$ to 330 $\Omega$	28 $\mu\Omega/\Omega$ + 2 m $\Omega$	
	330 $\Omega$ to 1.1 k $\Omega$	28 $\mu\Omega/\Omega$ + 2 m $\Omega$	
	1.1 k $\Omega$ to 3.3 k $\Omega$	28 $\mu\Omega/\Omega$ + 20 m $\Omega$	
	3.3 k $\Omega$ to 11 k $\Omega$	28 $\mu\Omega/\Omega$ + 20 m $\Omega$	
	11 k $\Omega$ to 33 k $\Omega$	28 $\mu\Omega/\Omega$ + 200 m $\Omega$	
	33 k $\Omega$ to 110 k $\Omega$	28 $\mu\Omega/\Omega$ + 200 m $\Omega$	
	110 k $\Omega$ to 330 k $\Omega$	32 $\mu\Omega/\Omega$ + 2 $\Omega$	
	330 k $\Omega$ to 1.1 M $\Omega$	32 $\mu\Omega/\Omega$ + 2 $\Omega$	
	1.1 M $\Omega$ to 3.3 M $\Omega$	60 $\mu\Omega/\Omega$ + 30 $\Omega$	
	3.3 M $\Omega$ to 11 M $\Omega$	130 $\mu\Omega/\Omega$ + 50 $\Omega$	
	11 M $\Omega$ to 33 M $\Omega$	250 $\mu\Omega/\Omega$ + 2.5 k $\Omega$	
	33 M $\Omega$ to 110 M $\Omega$	500 $\mu\Omega/\Omega$ + 3 k $\Omega$	
110 M $\Omega$ to 330 M $\Omega$	0.3 m $\Omega/\Omega$ + 100 k $\Omega$		
330 M $\Omega$ to 1 100 M $\Omega$	1.5 m $\Omega/\Omega$ + 500 k $\Omega$		



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Torque Tools - Wrenches, Drivers, Watches, Limiters, Multipliers <sup>FO</sup>	5 ozf*in to 100 ozf*in 5 lbf*in to 1 000 lbf*in 25 lbf*ft to 250 lbf*ft 250 lbf*ft to 600 lbf*ft	0.9% of Reading + 0.087 ozf*in 1.1% of Reading + 0.15 lbf*in 1% of Reading + 0.01 lbf*ft 0.61% of Reading + 0.73 lbf*ft	Torque Tester IC-ME-5

### Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure Pneumatic Pressure <sup>FO</sup>	Up to 29 psi 29 psi to 290 psi 290 psi to 2 900 psi	0.011 % + 0.0018 psi 0.0054 % + 0.013 psi 0.0057 % + 0.13 psi	Pressure Calibrator IC-TH-8
Equipment to Measure Hydraulic Pressure <sup>FO</sup>	Up to 5 000 psi 5 000 to 20 000 psi 20 000 to 40 000 psi	0.083 % + 0.56 psi 12 psi 25 psi	Pressure Calibrator IC-TH-8
Equipment to Measure Vacuum <sup>FO</sup>	-14.5 to 0 psi	0.002 psi	Pressure Calibrator IC-TH-8
Temperature Measuring Equipment <sup>FO</sup>	-196 °C to 0 °C 0 °C to 200 °C 200 °C to 1 000 °C	0.045 °C 0.035 °C 2.9 °C	Metrology Well/Bath/Furnace, PRT IC-TH-3
Infrared Thermometers <sup>FO</sup>	35 °C to 500 °C	0.38 % + 0.3 °C	Blackbody IC-TH-16
Equipment to Measure Humidity <sup>FO</sup>	10 % RH to 95 % RH	0.53 % RH	Thunder Scientific 2500 IC-TH-4

### Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Timers/Stopwatches <sup>FO</sup>	Up to 24 hrs	0.058 s/day	Timometer Time Base Method NIST 960-12



# Certificate of Accreditation: Supplement

## Industrial Calibration Service (In-Cal)

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force & Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Weight <sup>FO</sup>	Up to 2 kg	0.001 1 g	Mass Comparator Class 2 Std. Weights IC-ME-1
	2 kg to 5 kg	0.045 g	
	5 kg to 25 kg	0.29 g	
Analytical Balances <sup>FO</sup>	1 g to 5 000 g	0.000 56 % of Reading + 0.000 23 g	Ultra Class Weights IC-ME-3
Scales <sup>FO</sup>	Up to 600 lbf	0.014 % of Reading + 0.008 lbf	Class F Weights IC-ME-2
Equipment to Measure Force - Compression <sup>FO</sup>	10 lbf to 500 lbf	0.005 3 % of Reading + 0.034 lbf	Load Cell IC-ME-11
	500 lbf to 2 000 lbf	0.016 % of Reading	
	2 000 lbf to 800 klbf	0.5 % of Reading	
Equipment to Measure Force - Tension <sup>FO</sup>	10 lbf to 500 lbf	0.005 3 % of Reading + 0.034 lbf	
	500 lbf to 2 000 lbf	0.016 % of Reading	

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
- The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
- The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
- The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.



## *Certificate of Accreditation: Supplement*

### **Industrial Calibration Service (In-Cal)**

71 Pine Road, Hudson, NH 03051

Contact Name: Richie Fiore Phone: 603-883-5558

*Accreditation is granted to the facility to perform the following calibrations:*

5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
7. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.

